



Can the British Heart Foundation PocketCPR application improve the performance of chest compressions during bystander resuscitation: A randomised crossover manikin study.

Journal:	<i>Health Informatics Journal</i>
Manuscript ID	HIJ-15-0174.R1
Manuscript Type:	Original Article
Date Submitted by the Author:	01-Apr-2016
Complete List of Authors:	Eaton, Georgette; South Central Ambulance Service NHS Foundation Trust; Oxford Brookes University Renshaw, John; West Midlands Ambulance Service NHS Trust; Coventry University Gregory, Pete; University of Wolverhampton Kilner, Tim; University of Worcester
Keywords:	assistive technologies, PocketCPR, Bystander, mobile health, healthcare service innovation and IT
Abstract:	<p>This study aims to determine whether the British Heart Foundation (BHF) PocketCPR application can improve the depth and rate of chest compression, and therefore be confidently recommended for bystander use.</p> <p>118 candidates were recruited into a randomised crossover manikin trial. Each candidate performed CPR for two-minutes without instruction, or performed chest compressions using the PocketCPR application. Candidates then performed a further two minutes of CPR within the opposite arm.</p> <p>The number of chest compressions performed improved when PocketCPR was used compared to chest compressions when it was not (44.28% v 40.57, $P < 0.001$). The number of chest compressions performed to the required depth was higher in the PocketCPR group (90.86 v 66.26). The BHF PocketCPR application improved the percentage of chest compressions that were performed to the required depth. Despite this, more work is required in order to develop a feedback device that can improve bystander CPR without creating delay.</p>

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60



SCHOLARONE™
Manuscripts

For Peer Review

Title:

Can the British Heart Foundation PocketCPR application improve the performance of chest compressions during bystander resuscitation: A randomised crossover manikin study.

Word Count: 3,018

For Peer Review

1. Introduction

The provision of effective bystander cardiopulmonary resuscitation (CPR) in out-of-hospital cardiac arrest (OOHCA) remains unacceptably low¹⁻⁵ despite evidence that suggests that effective bystander CPR is associated with more favourable clinical outcomes and improved survival rates⁶. It is recognised that CPR is frequently inadequate when performed by laypersons^{1,7}, with many responders reluctant to perform mouth-to-mouth resuscitation during CPR⁸. To encourage uptake of bystander CPR attempts within the United Kingdom (UK), the British Heart Foundation⁸ launched a campaign to promote chest compression only CPR, with multiple studies supporting this approach⁹⁻¹². As part of their campaign, the British Heart Foundation produced a smartphone PocketCPR training application to provide real-time feedback on the depth of chest compressions performed during CPR, and provide metronomic feedback to ensure accurate external chest compression rate. The importance of adequate chest compressions depth, and accurate rate of compressions were both reaffirmed within current resuscitation guidelines¹³, with suboptimal compression rates associated with poor return of spontaneous circulation¹⁴. Nevertheless, the performance of both compression depth and compression rate by bystanders is shown to be suboptimal¹⁵⁻¹⁶.

To improve the performance of chest compressions, feedback systems have been used successfully in training to improve the overall quality of layperson CPR^{3,4,6,9,17} and maintain skill acquisition and retention¹⁷⁻¹⁸, although there has been insufficient evidence to validate these applications for use in practice. This study endeavours to determine whether the British Heart Foundation PocketCPR feedback application would improve chest compression performance during bystander resuscitation.

2. Materials and Methods

2.1 Objectives

The aim of this randomised crossover manikin study was to investigate whether using the British Heart Foundation PocketCPR training application would improve the performance of chest compressions against current resuscitation guidelines when used by laypersons with no recent CPR training.

We hypothesised that the British Heart Foundation PocketCPR training application would increase the proportion of chest compressions performed at the recommended depth of 50 to 60mm, and improve the rate of chest compressions per minute, compared to no application.

2.2 Participants and randomisation

Participants were recruited from a University campus on an opportunistic basis. All participants were required to be aged 18 years or over; not be a healthcare professional; and not having attended a CPR training course in the last six months. This last point was pertinent, since it is acknowledged in the literature that skills and knowledge relating to bystander CPR decay rapidly after initial training^{19, 20, 21}. Volunteers were provided with a participant information sheet and an opportunity to ask questions of the researchers before being asked to give written consent.

2.3 Methodology

1
2
3 In this randomised crossover study, each participant was asked to attempt a two
4
5 minute period of CPR on a Laerdal Resuscitation manikin (Resusci Anne Skills
6
7 Station, Laerdal Medical Limited, Orpington, UK) with and without the PocketCPR
8
9 application in accordance with a pre-randomised order. Candidates were not
10
11 required to have previous experience or ownership of a smartphone device to take
12
13 part and were provided with information as to how to hold the iPod and activate the
14
15 PocketCPR software. Randomising the participants reduced the risk that the
16
17 participant may perform better in the second arm of the study where they had
18
19 previously used PocketCPR. The randomisation order was generated using the
20
21 statistical software package PASW (version 17.0.2, SPSS Inc, Chicago, Illinois,
22
23 USA) and ensured that 50% of the participants performed first with the PocketCPR
24
25 application and 50% without. The PocketCPR application was used on an iPod
26
27 Touch 2009 device.
28
29
30

31
32 Participants were required to rest for two minutes between each CPR scenario to
33
34 ensure that operator fatigue did not adversely affect the second attempt. The
35
36 instructions given to participants were limited to information on how to hold the iPod
37
38 and how to activate the British Heart Foundation PocketCPR software.
39
40

41
42 The software gives visual feedback in the form of a bar on the display indicating
43
44 current compression depth with a green colour marking the ideal interval, and verbal
45
46 feedback prompts (including “press harder”, “press faster”, “press slower”, “good
47
48 depth”) (Figure 1). Additionally, the device has an integrated metronome which ticks
49
50 at a rate of 100 per minute signalling the correct compression rate. There was no
51
52 feedback on ventilations since this application is designed to support chest
53
54 compression only CPR. When performing CPR without the device, the participant
55
56 received no verbal or visual feedback and no metronome guidance
57
58
59
60

2.4 Data collection

Performance measurements derived from the manikin software were recorded onto a connected laptop. These were compression rate; compression depth; hand position for performing chest compressions; tidal volume of ventilation attempt; and time off the chest once CPR had been started. Additional observations (including time to start CPR) were recorded manually by the researchers. Manual observations included any ventilation attempt and rate which did not register in the manikin due to an occluded airway.

2.5 Statistical analysis

Sample size was calculated based on the primary outcome measure of the adequate depth of compression. Previous research has reported that 42% of trained prehospital providers delivered chest compressions with a mean depth of 50-60mm during one minute simulated cardiac arrest scenario, compared with 39% listening to a musical prompt (correlation coefficient $[\phi] = 0.44051$)²². In order to detect 15% (from 42% to 57%) increase in the proportion of lay persons delivering compressions at the recommended depth with a power of 0.85 and an alpha of 0.05, it was estimated that 108 subjects were required (sample size for paired cohort study calculated using StatsDirect, version 2.7.8 StatsDirectLtd, Altrincham UK).

Analysis compared the difference in performance of chest compressions with and without the British Heart Association PocketCPR application. The primary outcome

1
2
3 measure of mean compression depth was analysed alongside secondary outcome
4
5 measures of mean total compressions in 2 minutes of simulated CPR; mean
6
7 compression rate; mean total correct compressions; and correct hand position.
8
9

10 The quality of chest compressions was measured with Resusci Anne Skills Station
11 (Laerdal Medical Limited, Orpington, UK). IBM SPSS Statistics version 22 software
12 package was used to calculate descriptive statistics, p values, 95% confidence
13 intervals and Wilcoxon's rank sum 2 related samples. A significance level of $p < 0.05$
14
15 was considered to be statistically significant.
16
17
18
19
20
21

22 *2.6 Ethical Considerations*

23
24
25
26

27 This randomised control trial received ethical approval from the Coventry University
28 Ethics Committee (P4090).
29
30
31
32

33 **3. Results**

34
35
36

37 *3.1 Flow and baseline characteristics*

38
39

40 118 subjects were recruited to the study and were included in the analysis. The
41 sample size of 108 subjects determined by the power sample size calculation was
42 satisfied. Baseline characteristics are shown in Figure 2.
43
44
45
46

47 *3.2 Primary Outcome*

48
49

50 When using the PocketCPR application, 44.28% of the total number of compressions
51 were measured to be the correct depth compared with 40.57% of mean total
52 compressions when the PocketCPR app was not used ($p < 0.001$). The actual number
53 of correct depth compressions was also higher in the PocketCPR group than would
54
55
56
57
58
59
60

1
2
3 be anticipated by the percentages (90.86 v 66.24) as this group performed more
4
5 compressions in the two minute period.
6
7

8 *3.3 Secondary Outcomes*

9

10
11 Further analysis was performed on the secondary outcome measures and the results
12
13 reported in table 2.
14

15
16 The 2010 Resuscitation Council (UK) guidelines¹³ advocate a compression rate of
17
18 100-120 compressions per minute, therefore continuous chest compressions over a
19
20 2 minute period should result in 200-240 compressions being delivered over the 2
21
22 minute period. Whilst the mean compression rate when using the PocketCPR was
23
24 broadly similar to the mean compression rate when not using the application, there
25
26 was a significant difference in the total number of compressions performed. When
27
28 using the PocketCPR application, the number of compressions delivered fell within
29
30 the expected range whilst the number in the non-PocketCPR group was lower
31
32 (p<0.000)
33
34
35
36

37
38 There was no significant difference between mean compression rates, but there was
39
40 a significant difference between the number of compressions performed during the 2
41
42 minute time period.
43

44
45 The difference in the mean number of correct compressions when using PocketCPR
46
47 and without the application was not significant, and the mean number of correct
48
49 compressions was low in both groups. 14.94% of the mean total compressions (95%
50
51 CI 20.38-40.96) were correct in terms of rate, depth and hand position when using
52
53 the application, 12.55% of the mean total compressions (95% CI 13.03-27.97)
54
55 without the application.
56
57
58
59
60

1
2
3 Where compressions were incorrect due to incorrect hand placement, hands were
4 more likely to be too low rather than too high. Where hands were placed away from
5 the midline, subjects were most likely to place their hands further from their body to
6 the left of the mid line of the manikin, rather than closer to them to the right of the
7 midline of the manikin.
8
9
10
11
12

13 14 15 **4. Discussion**

16 17 18 *4.1 Comment*

19
20 The PocketCPR training application allowed for a greater depth of compressions
21 during the two minute resuscitation attempt, which supports the recommendations
22 for CPR in current guidelines¹³. Participants were more likely to reach the
23 recommended resuscitation guidance depth of 50-60mm for chest compressions
24 when using the PocketCPR application (CI 81.19-112.82, P<0.006). With evidence
25 connecting an increased likelihood of return of spontaneous circulation with chest
26 compressions performed to a depth of 50mm or more²³⁻²⁴, the results here suggest
27 that the feedback provided by the PocketCPR application could have real-life
28 applications. Previous manikin based studies have found that compressions tend
29 towards inadequate depth following only one minute of CPR due to rescuer fatigue²⁵,
30 whereas our results demonstrate that feedback to the layperson allowed chest
31 compressions to be performed to an adequate depth more frequently during a two
32 minute cycle. This suggests that the PocketCPR application may either ameliorate
33 the effect of rescuer fatigue or help to motivate the rescuer to continue to compress
34 the chest to an adequate depth even when fatigue or a loss of concentration is taking
35 effect.
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56

57 Whilst the depth of chest compressions improved with the PocketCPR application, it
58
59
60

1
2
3 was observed that hand position was frequently reported off centre (Figure 3). This
4
5 study found that incorrect hand position was often too high on the chest, rather than
6
7 too low or too far left or right. As a percentage, there was no difference in hand
8
9 positioning between participants using the PocketCPR device or those without. The
10
11 use of PocketCPR likely required more dexterity to hold the device between the
12
13 hands during chest compressions as well as the requirement to visualise the screen
14
15 to view depth attainment, although our results failed to reach significance. There is
16
17 insufficient evidence in research to determine if there is any relationship between
18
19 incorrect hand position and changes in efficiency of CPR²⁵. It may also be
20
21 considered that CPR performed too right or too left is more likely to be effective than
22
23 chest compressions performed too high and too low due to changes in thoracic
24
25 pressure as part of the thoracic pump theory²⁶.
26
27
28
29

30
31 Despite the high number of incorrect hand positions in both groups, use of
32
33 PocketCPR did result in a greater number of chest compressions that were
34
35 performed during the two-minute resuscitation attempt. When using the PocketCPR
36
37 application, participants achieved the predicted range of 200-240 chest
38
39 compressions, over the 2 minute test period, in accordance with resuscitation
40
41 guidance of 100-120 chest compressions per minute. Although the number of chest
42
43 compressions performed using PocketCPR was higher than without, there was a
44
45 noticeable delay in starting chest compressions whilst participants navigated the
46
47 British Heart Foundation PocketCPR training application. Whilst any lack of
48
49 familiarity with the device was ameliorated by the instructions that were given to the
50
51 participants, on average chest compressions were delayed by 37.31 seconds while
52
53 candidate's summoned help, confirmed that the patient was not breathing, and
54
55 began chest compressions with device feedback. The delay to begin chest
56
57
58
59
60

1
2
3 compression is consistent with comparable studies and similar portable feedback
4
5 devices²⁵. Delay to CPR without the device was on average 14.42 seconds. The
6
7 PocketCPR device appears to improve the consistency of chest compressions, as
8
9 participants still performed a greater number of chest compressions in the required
10
11 time, albeit with an initial delay. It is unclear whether the delay to start initial chest
12
13 compressions is countered by the increased number of compressions achieved over
14
15 two minutes but the delay in starting could be further exacerbated where the
16
17 application is not readily available on the home screen of an individual about to
18
19 perform bystander CPR. A recommendation would be to ensure that the application
20
21 is pinned to a person's home page on their mobile device in order to minimise any
22
23 delays in starting chest compressions. To overcome this delay in beginning chest
24
25 compressions, it is recommended that instruction is more concise, bypassing the
26
27 approach and navigating straight to resuscitation feedback. Without the application,
28
29 CPR was often commenced earlier, but had lower consistency and longer periods of
30
31 inactivity due to some participants attempting mouth-to-mouth ventilation. Since
32
33 bystander resuscitation with periods of inactivity is associated with poorer outcome²⁷,
34
35 it is important to minimise this inactivity. In our study, those performing chest
36
37 compression only CPR using the PocketCPR application had fewer periods of
38
39 inactivity and more consistency in their compressions. The periods of inactivity serve
40
41 to explain why the number of compressions in the PocketCPR arm of the study
42
43 performed significantly more compressions during the two minutes of the study yet
44
45 there was no significant difference in the actual rate of compressions between the
46
47 two limbs.
48
49
50
51
52

53
54
55 During both limbs of the study subjects managed to achieve an average rate of
56
57 compression that accords with current resuscitation guidelines, which suggest that
58
59
60

1
2
3 chest compressions of at least 100 compressions per minute are more effective than
4
5 slower rates²⁸.
6
7

8 Participant recruitment was representative of a normal urban population. Since
9
10 OHCA can occur in any environment, the participant demographics are
11
12 representative of those who may render aid in this situation. During the study it was
13
14 noted that most participants were familiar with the device and navigation through the
15
16 application, but were more often unable to effectively follow the instructions as
17
18 directed. Although we collected information on age, there are many factors that
19
20 affect the rescuers ability or willingness to perform bystander CPR. These include
21
22 (but are not limited to) socioeconomic profile; education²⁹; gender and fitness³⁰.
23
24
25
26
27

28 Participants in this study were seen to encounter navigation problems whilst the
29
30 application was playing, due to the touch-screen nature of the device. This resulted
31
32 in accidental disruption of CPR instruction and restarting the application, which
33
34 adversely impacted upon the time to perform chest compressions. It would be helpful
35
36 if the device became locked once the application had been selected and the
37
38 accelerometers activated by chest compressions. Despite this, the benefit of using
39
40 the PocketCPR application is that participants only required a smartphone device
41
42 with the application, rather than additional equipment to secure the device. Whilst
43
44 other studies have shown improved CPR with the use of smartphones secured in
45
46 armbands, other studies demonstrate that participants were unable to perform CPR
47
48 with feedback without the securing mount³¹. Whilst the PocketCPR uses smartphone
49
50 technology to feedback depth measurement, gripping the device whilst performing
51
52 chest compressions is difficult and the usability is compromised due to the accidental
53
54 disruption of CPR instruction.
55
56
57
58
59
60

1
2
3
4
5
6 There is a growing body of evidence to suggest that using a smartphone application
7
8 can improve chest compression depth and rate³², which positively affects chance of
9
10 survival in OHCA. However, these devices are limited in their usability and complex
11
12 interfaces, which must be overcome to confidently recommend their use beyond a
13
14 training application into a real-time feedback device. Smartphone applications may
15
16 also be useful as prerequisite learning for CPR training³³.
17
18
19

20
21 Although this study considers PocketCPR application to improve chest compression
22
23 performance, the evolution of portable smart technology is becoming ever more
24
25 prevalent in prehospital resuscitation, evident by the recent endorsement of the good
26
27 Smartphone Activation Medics (GoodSAM) application from the Resuscitation
28
29 Council UK³⁴, which alerts nearby rescuers to cases of OOHCA. Therefore, there is
30
31 opportunity to develop applications like PocketCPR to combine rescuer activation
32
33 with effective CPR feedback until professional help arrives.
34
35
36
37
38

39 *4.2 Limitations*

40
41
42 There were several limitations within this study. Firstly, over compression of chest
43
44 compressions was not measurable on the resuscitation manikin as the physical
45
46 design of the manikin prevented the participant from over-compressing. Data were
47
48 collected to consider whether compression depth was insufficient, but it is not known
49
50 how many compressions may have been too deep. Despite this, there is insufficient
51
52 evidence to specify an upper limit for chest compression depth, and chest
53
54 compressions that are too deep may still be effective³⁵.
55
56
57
58
59
60

1
2
3 Secondly, the number of people who have PocketCPR application downloaded onto
4 their portable device or smartphone and accessible during OHCA incidents may limit
5 the usability of feedback. It is not known how many times the application has been
6 downloaded but it works on both an Apple and Android platforms so there is
7 considerable potential for this application be widely available. In the first quarter of
8 2015 alone, over 74 million iPhones were sold as a standalone mobile product ³⁶. It
9 is not unreasonable to argue that the application should be included as a default
10 application on all devices capable of supporting it.
11
12

13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Lastly, as with all simulation and manikin studies, the results of this study cannot measure clinical outcome or survivability, but remains a useful proxy measure into the usability of feedback devices for bystanders.

5. Conclusion

Overall, the standard of bystander resuscitation within this study was poor and chest compressions were still frequently performed at insufficient depth, with incorrect hand positioning and with prolonged periods of inactivity. The PocketCPR application improved the percentage of chest compressions that were performed to the correct depth during bystander compression-only CPR. A greater number of chest compressions was also performed with the application during the two-minute time period when compared to standard basic life support attempts, where compressions were often too shallow and with too few external chest compressions performed. Although use of the application improved CPR performance when compared to no application, CPR performance remained suboptimal. More work is needed to develop an application that can instruct bystanders to perform effective chest compression only CPR without delay.

Conflicts of interest Statement

None of the authors have any conflict of interest with devices, organisations or applications mentioned in this article, nor with any competitors.

Acknowledgement

The authors would like to thank Malcolm Woollard for his support in the early stages of this study as well as Coventry University for their facilitation.

References

1. Spooner B, Fallaha F, Kocierz L, Smith C, Smith S, Perkins G. An evaluation of objective feedback in basic life support (BLS) training. *Resuscitation* 2007;73:417-24
2. Iwami T, Kawamura T, Hiraide A, et al. Effectiveness of bystander-initiated cardiac-only resuscitation for patients with out-of-hospital cardiac arrest. *Circulation* 2007;116:2900-07
3. Sayre R, Berg A, Cave M, Page R, Potts J, White R. Hands-only (compression-only) cardiopulmonary resuscitation: A call to action for bystander response to adults who experience out-of-hospital sudden cardiac arrest: a science advisory for the public from the American Heart Association Emergency Cardiovascular Care Committee. *Circulation* 2008;117:2162-67

- 1
2
3 4. Drager K. Improving patient outcomes with compression only CPR: Will
4 bystander CPR rates improve? *J Emerg Nurs* 2012;38: 243-48
- 5
6
7 5. Takei Y, Nishi T, Matsubara H, Hashimoto M, Inaba H. Factors associated with
8 quality of bystander CPR: The presence of multiple rescuers and bystander-
9 initiated CPR without instruction. *Resuscitation* 2014;85:492-98
- 10
11
12 6. Zanner R, Wilhelm D, Feussner H, Schneider G. Evaluation of M-AID, a first
13 aid application for mobile phones. *Resuscitation* 2007; 74:487-94
- 14
15
16 7. Aufderheide P, Pirralo G, Yannopoulos D, et al. Incomplete chest wall
17 decompression: A clinical evaluation of CPR performance by trained
18 laypersons and an assessment of alternative of manual chest compression-
19 decompression techniques. *Resuscitation* 2006;71: 341-51
- 20
21
22 8. British Heart Foundation. Hands only CPR FAQs. London, British Heart
23 Foundation (2013) (Accessed 30 March 2015 at [http://www.bhf.org.uk/heart-](http://www.bhf.org.uk/heart-health/life-saving-skills/hands-only-cpr-faqs.aspx)
24 health/life-saving-skills/hands-only-cpr-faqs.aspx)
- 25
26
27 9. Urban J, Thode H, Stapleton E. Current knowledge of and willingness to
28 perform Hands-Only™ CPR in laypersons. *Resuscitation* 2013; 84:1574-1578
- 29
30
31 10. Waalewijn RA, Tijssen JGP, Koster RW. Bystander initiated actions in out-of-
32 hospital cardiopulmonary resuscitation: Results from the Amsterdam
33 resuscitation study (arrest). *Resuscitation* 2001;50:273–9
- 34
35
36 11. Bobrow BJ, Spaite DW, Berg RA et al. Chest compression–only CPR by lay
37 rescuers and survival from out-of-hospital cardiac arrest. *JAMA*
38 2010;304:1447–54.
- 39
40
41 12. Nagao K, Kikushima K, Sakamoto T, et al. Cardiopulmonary resuscitation by
42 bystanders with chest compression only (SOS-KANTO): An observational
43 study. *Lancet* 2007;369:920
- 44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

- 1
2
3 13. Nolan J, Soar J, Zideman D, et al. European Resuscitation Council Guidelines
4 for Resuscitation 2010 Section 1. Executive summary. Resuscitation
5 2010;81:1219-1276
6
7
8
9
10 14. Abella B, Sandbo N, Vassilatos P, et al. Chest compression rate during
11 cardiopulmonary resuscitation are suboptimal: a prospective study during in-
12 hospital cardiac arrest. Circulation 2005;111:428-38
13
14
15
16 15. Nishiyama C, Iwami T, Kawamura T et al. Effectiveness of simplified chest
17 compression-only CPR training for the general public: A randomised controlled
18 trial. Resuscitation 2008;79:90-96
19
20
21
22
23 16. Birkenes T, Myklebust H, Neset A, Kramer-Johansen J. Quality of CPR
24 performed by trained bystanders with optimised pre-arrival instructions.
25 Resuscitation 2014;85: 124-130
26
27
28
29
30 17. Yeung J, Meeks R, Edelson D, Gao F, Soar J, Perkins G. The use of CPR
31 feedback/prompt devices during training and CPR performance: A systematic
32 review. Resuscitation 2009;80:743-751
33
34
35
36
37 18. Low D, Clark N, Soar J, Padkin A, Stoneham A, Perkins G, Nolan J. A
38 randomised controlled trial to determine if the use of iResus application on a
39 smart phone improves the performance of an advanced life support provider in
40 a simulated medical emergency. Anaesthesia 2011;66: 255-62
41
42
43
44
45 19. Woollard M, Whitfield R, Smith A, et al. Skill acquisition and retention in
46 automated external defibrillator (AED) use and CPR lay responders: a
47 prospective study. Resuscitation 2004; 60:17-28
48
49
50
51
52 20. Creutzfeldt J, Hedman L, Medin C, Stengård K, Felländer-Tsai L. Retention of
53 knowledge after repeated virtual world CPR training in high school students.
54 Stud Health Technol Inform 2009;142: 59-61.
55
56
57
58
59
60

- 1
2
3 21. Isbye D, Meyhoff C, Lippert F, Rasmussen L. Skill retention in adults and
4 children 3 months after basic life support training using a simple personal
5 resuscitation manikin. *Resuscitation* 2007;74: 296-302.
6
7
8
9
10 22. Woollard M, Poposki J, McWhinnie B, Rawlins L, Munro G, O'Meara P. Achy
11 Breaky Makey Wakey Heart? A randomised cross-over trial. *Emerg Med J*
12 2011 (Accessed 30 March 2015, at
13 <http://emj.bmj.com/content/early/2011/10/19/emmermed-2011-200187.full>)
14
15
16
17
18 23. Rajab T, Pozner C, Conrad C, Cohn L, Schmitto J. Technique for chest
19 compressions in adult CPR. *World J Emerg Surg* 2011;41. (Accessed 7 June
20 2015, at [http://www.ncbi.nlm.nih.gov/pmc/articles/PMC3261806/pdf/1749-](http://www.ncbi.nlm.nih.gov/pmc/articles/PMC3261806/pdf/1749-7922-6-41.pdf)
21 [7922-6-41.pdf](http://www.ncbi.nlm.nih.gov/pmc/articles/PMC3261806/pdf/1749-7922-6-41.pdf))
22
23
24
25
26
27 24. Hightower D, Thomas S, Stone C, Dunn K, March J. Decay in quality of
28 closed-chest compressions over time. *Ann Emerg Med*. 1995;26: 300 –303
29
30
31 25. Zapletal B, Greif R, Stumpf D, et al. Comparing three CPR feedback devices
32 and standard BLS in a single rescuer scenario: a randomised simulation study.
33 *Resuscitation* 2014; 85; 560-566
34
35
36
37 26. Chul Cha K, Jun Kim Y, Jin Shin H, et al. Optimal position for external chest
38 compression during cardiopulmonary resuscitation: an analysis based on chest
39 CT in patients resuscitated from cardiac arrest. *Emerg Med J* 2013;30: 615-
40 619
41
42
43
44
45
46
47 27. Eftestøl T, Sunde K, Steen P. Effect of interrupting precordial compressions
48 on calculated probability of defibrillation success during out-of-hospital cardiac
49 arrest. *Circulation* 2002; 105; 2270-2273
50
51
52
53
54 28. Nolan J, Perkins G, Soar J. Chest compression rate: Where is the sweet spot?
55 *Circulation* 2012; 125; 2968-2970
56
57
58
59
60

- 1
2
3 29. Papalexopoulou K, Chalkias A, Dontas I, et al. Education and age affect skill
4 acquisition and retention in lay rescuers after a European Resuscitation
5 Council CPR/AED course. *Heart Lung*. 2014;43:66-71
6
7
8
9
10 30. Reddy K, Murray B, Rudy S, et al. Abstract 224: Effective Chest Compressions
11 Are Related to Gender and Body Mass Index. *Circulation*. 2011;124: A224
12
13
14 31. Semeraro F, Taggi F, Tammaro G, et al. iCPR: A new application of high-
15 quality cardiopulmonary resuscitation training. *Resuscitation* 2010;82:436-441
16
17
18 32. Sakai T, Kitamura T, Nishiyama C, et al. Cardiopulmonary resuscitation
19 support application on a smartphone – randomized controlled trial. *Circ J*.
20 2015;79: 1052-1057
21
22
23
24 33. Park S. Comparison of chest compression quality between the modified chest
25 compression method with the use of smartphone application and the
26 standardized traditional chest compression method during CPR. *Technol*
27 *Health Care*. 2014;22: 351-358
28
29
30
31
32
33
34 34. Resuscitation Council UK. The RC (UK) endorses the GoodSAM app.
35 <https://www.resus.org.uk/statements/the-rc-uk-endorses-the-goodsam-app/>
36 (2016, accessed 7 March 2016)
37
38
39
40 35. Stiell I, Brown S, Christenson J, et al. What is the Role of Chest Compression
41 Depth during Out-of- Hospital Cardiac Arrest Resuscitation? *Crit Care Med*.
42 2012;40:1192–1198
43
44
45
46
47 36. Apple inc. Press Releases. January 2015: Apple Reports Record First Quarter
48 Results. [http://www.apple.com/uk/pr/library/2015/01/27Apple-Reports-Record-](http://www.apple.com/uk/pr/library/2015/01/27Apple-Reports-Record-First-Quarter-Results.html)
49 [First-Quarter-Results.html](http://www.apple.com/uk/pr/library/2015/01/27Apple-Reports-Record-First-Quarter-Results.html) (2015, accessed 25 June 2015)
50
51
52
53
54
55
56
57
58
59
60

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60




Figure 1. Screenprint of PocketCPR feedback screen.
129x71mm (96 x 96 DPI)

Peer Review

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Figure 2. Demographic Table of participants (n=120)	
Gender	Total
Male	58
Female	62
Age Group (years)	Total
18-25	68
26-33	11
34-41	8
41-48	22
49+	10

Table 1. Demographic Table of Participants
91x106mm (96 x 96 DPI)

Review

Parameter	With App Mean %	95%CI	Without app Mean %	95% CI	p
<i>Predicted max compressions</i>	200-240		200-240		
Total compressions	205.19	199.11-211.24	163.25	151.24-175.25	.000
Comp rate	106.87	104.87-108.88	105.37	100.58-110.17	.858
Correct compressions	30.67 (14.94)	20.38-40.96	20.5 (12.55)	13.03-27.97	.085
Adequate depth	90.86 (44.28)	75.74-105.99	66.24 (40.57)	52.5-79.97	.001
Insufficient depth	114.32 (55.71)	99.32-129.33	97.01 (59.42)	81.19-112.82	.006
Low hand position	48.1 (23.44)	34.54-61.66	44.97 (27.54)	32.68-57.26	.970
High hand position	32.52 (15.84)	21.96-43.08	28.29 (17.32)	18.49-38.09	.351
Right hand position	11.42 (5.56)	3.97-18.86	4.41 (2.7)	-0.2-9.02	.194
Left hand position	28.36 (13.82)	17.6-39.12	23.77 (14.56)	14.21-33.33	.788

Table 2. Results Table
318x99mm (96 x 96 DPI)